Model: RGT9-60STD-HP3

Certificate Number: 3109

1page / Total3page Measurement Date:June 5,2024 Certificate Issue date:June 11,2024

Measurement Certificate

Requester RAYGEARS LLC

Address 7-4-15-4F Honcho, Funabashi-City, Chiba Prefecture, Japan

Product Name TSUKUYOMI • 9(VR.3)

Model RGT9-60STD-HP3

Manufacturer RAYGEARS LLC

Measurement Unit Illuminance [lx]

Measurement Method OTCL Calibration procedure(LAB-T001), ANSI / NEMA FL1

Environment of Measurement Temperature: 23 °C ± 2 °C Relative Humidity: 65 % ± 20 %

Measurement Date June 5, 2024

Certificate Issue Date June 11, 2024

Kyokko Trading Co.,Ltd.

Optical Test & Calibration Laboratory

Optical Metrology Engineer

Takuro Hirano



旭光通商株式会社 光学試験校正室

光学試験校正室(ISO/IEC 17025 認定校正機関) 〒105-0014 東京都港区芝1-14-4 芝桝田ビルB1F TEL: 03 6371-6908 FAX: 03-63/1-6944 Optical Test and Calibration Laboratory
Shibamasuda Bldg. B1F, 1-14-4 Shiba, Minato-ku, Tokyo
105-0014 JAPAN Phone: +81-3-6371-6908 Fax: +81-3-6371-6944

Takuro Hu

This company that issued this certificate complies with JIS Q 17025(ISO/IEC 17025:2017), and it is accredited by JAB, which is a member of ILAC and APAC-MRA. Do not reproduction and use of only a portion of this certificate without the consent of the issuing authority. The measurement values in this certificate are traceable to the SI unit system.

Model: RGT9-60STD-HP3

Certificate Number: 3109

Mearsurement Results

1.Measurement Content


1-1: Illuminance Measurement Results

Illuminance[klx]	RGT9-60STD-HP3		
Measurement Value	123.2		

1-2: Measurement Envoirment

Temperature: 21.2 °C Relative Humidity: 45.4%

- 1. A measurement distance is 2.0 m between B520 illuminance meter's photometer head surface and measuring sample's surface.
- 2. A sample is lit by the built-in battery.
- 3. The measurement was performed after lighting.
- 4. A measurement position is adjusted so that the illuminance meter's value was maximized.



1-3: Reference Standards used for calibration

Reference standards	Manufacturer	Model	S/N	Lamp used for	Calib.
NIST compliante Spectral Irradiance Working Standard	Optronic Laboratory	OL345RP	90101125		
Illuminance meter	LMT	B520	04A5181 04A5182	V	

Model: RGT9-60STD-HP3

Certificate Number: 3109

2.PEAK BEAM INTENSITY[cd]

2-1: PEAK BEAM INTENSITY[cd] Calclation

ANSI/NEMA FL1 Section 2.3.6, obtain the illuminance results and "PEAK BEAM INTENSITY" from distance bytge the following formula(1).

Illuminance[$|x| \times (Distance[m])^2 = PEAK BEAM INTENSITY[cd] -----(1)$

* Illuminance[lx] and Distance[m] from Illuminance measuring results(1-1).

2-2: PEAK BEAM INTENSITY calcurate results

PEAK BEAM INTENSITY[cd]
492,800

3.BEAM DISTANCE[m]

3-1:BEAM DISTANCE[m] calcuration method ANSI/NEMA FL1 Section 2.2.6, obtain "BEAM DISTANCE" from "PEAK BEAM INTENSITY" by the following formula (2).

$$\sqrt{\text{(PEAK BEAM INTENSITY/ 0.25)}}$$
 = BEAM DISTANCE -----(2)

3-2:BEAM DISTANCE[m] calcuration results

BEAM DISTANCE[m]

1404